
Midmark M11 - v1.0.0
Total Cycles: 2798

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2023 12:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.8 F 0.0
2:00 211.5 F 4.2
4:00 217.1 F 9.6
6:00 223.5 F 6.2
8:00 233.3 F 10.5
10:00 234.1 F 7.6
12:00 242.0 F 11.2
14:00 246.4 F 13.4
16:00 259.2 F 20.4
18:00 270.4 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.4
1:00 272.6 F 29.1
2:00 271.8 F 28.3
3:00 272.6 F 29.1
4:00 271.8 F 28.4

Min 271.5 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.9 F 0.0
5:00 229.7 F 0.0
10:00 235.9 F 0.0
15:00 230.3 F 0.0
20:00 225.3 F 0.0
25:00 225.8 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 18:41
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:49
CYCLE COMPLETE

01/19/2023 01:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____