
Midmark M11 - v1.0.5
Total Cycles: 3189

Name: RITTER2
Sterilizer ID: V2313228
12/06/2023 11:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.6 F 0.0
2:00 148.1 F 0.9
4:00 161.6 F 3.5
6:00 183.7 F 7.4
8:00 203.9 F 11.8
10:00 224.7 F 6.0
12:00 229.4 F 10.9
14:00 231.8 F 7.6
16:00 244.6 F 13.0
18:00 244.2 F 12.6
20:00 256.3 F 19.0
22:00 267.1 F 25.8
24:00 272.9 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 271.6 F 29.5
2:00 271.4 F 29.6
3:00 271.5 F 29.5
4:00 271.6 F 29.4

Min 271.4 F 29.4
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.1 F 0.0
5:00 239.1 F 0.0
10:00 190.8 F 0.0
15:00 189.2 F 0.0
20:00 186.8 F 0.0
25:00 184.7 F 0.0
30:00 178.5 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:12
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:02:43
CYCLE COMPLETE

12/06/2023 12:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____