
Midmark M11 - v1.0.0
Total Cycles: 5031

Name: RITTER 1
Sterilizer ID: V2178198
10/29/2024 08:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.9 F 0.0
2:00 72.9 F 0.0
4:00 92.7 F 0.0
6:00 131.0 F 0.2
8:00 159.7 F 0.5
10:00 177.1 F 2.2
12:00 195.7 F 5.1
14:00 209.1 F 8.7
16:00 219.7 F 12.3
18:00 222.3 F 5.7
20:00 233.5 F 10.1
22:00 236.2 F 8.7
24:00 242.8 F 11.9
26:00 239.9 F 10.0
28:00 252.8 F 16.3
30:00 263.7 F 22.7
32:00 272.0 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.6 F 28.5
2:00 271.4 F 27.8
3:00 272.7 F 28.6
4:00 271.7 F 27.9

Min 271.4 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 204.7 F 0.0
10:00 209.6 F 0.0
15:00 217.3 F 0.0
20:00 227.4 F 0.0
25:00 222.9 F 0.0
30:00 218.2 F 0.0

Drying Complete

FILLING: 0:43
HEATING: 32:11
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:10:18
CYCLE COMPLETE

10/29/2024 09:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____