
Midmark M11 - v1.0.5
Total Cycles: 4454

Name: RITTER2
Sterilizer ID: V2313228
03/25/2025 02:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.9 F 0.0
2:00 124.7 F 1.2
4:00 138.4 F 3.8
6:00 170.2 F 7.6
8:00 193.1 F 11.2
10:00 212.4 F 4.3
12:00 220.9 F 8.7
14:00 229.8 F 13.1
16:00 234.2 F 10.0
18:00 236.9 F 9.3
20:00 245.8 F 14.1
22:00 257.3 F 20.3
24:00 267.7 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 30.1
1:00 271.3 F 30.0
2:00 271.5 F 30.0
3:00 271.8 F 30.2
4:00 271.8 F 30.5

Min 271.3 F 30.0
Max 272.4 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 234.4 F 0.0
10:00 226.1 F 0.0
15:00 222.5 F 0.0
20:00 206.3 F 0.0
25:00 204.6 F 0.0
30:00 199.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:23
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:04:19
CYCLE COMPLETE

03/25/2025 03:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____