
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000119
9:48 11/03/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:21
PRESSURIZING 1:27
STERILIZING 2:48
136.0 C 220kPa 2:48
136.3 C 223kPa 3:18
136.3 C 221kPa 3:48
136.1 C 221kPa 4:18
136.2 C 219kPa 4:48
136.3 C 222kPa 5:18
136.3 C 222kPa 5:48
136.4 C 222kPa 6:18
136.4 C 222kPa 6:18
VENTING 6:18
AIR DRYING 7:14
CYCLE COMPLETE 7:21

Digital Signature #
726CA49CAA603ECD
