

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000206  
8:48 08/07/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:05  
PRESSURIZING 3:09  
STERILIZING 4:10  
136.4 C 223kPa 4:10  
136.4 C 223kPa 4:40  
136.4 C 222kPa 5:10  
136.3 C 223kPa 5:40  
136.2 C 222kPa 6:10  
136.3 C 222kPa 6:40  
136.4 C 222kPa 7:10  
136.3 C 222kPa 7:40  
136.3 C 222kPa 7:40  
VENTING 7:40  
AIR DRYING 8:30  
CYCLE COMPLETE 8:33

-----  
Digital Signature #  
A1286530676D47C6  
-----