
Midmark M11 - v1.0.0
Total Cycles: 2469

Name: RITTER 1
Sterilizer ID: V2178198
10/21/2022 10:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.2 F 0.0
2:00 193.6 F 3.9
4:00 206.7 F 8.4
6:00 214.4 F 13.4
8:00 228.6 F 7.7
10:00 237.9 F 12.9
12:00 237.3 F 10.1
14:00 238.0 F 8.2
16:00 249.0 F 14.9
18:00 261.2 F 21.7
20:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.6 F 29.1
2:00 271.6 F 28.4
3:00 272.4 F 29.2
4:00 272.3 F 28.9

Min 271.4 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 235.8 F 0.0
10:00 233.5 F 0.0
15:00 230.7 F 0.0
20:00 228.9 F 0.0
25:00 229.0 F 0.0
30:00 230.3 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:55
CYCLE COMPLETE

10/21/2022 11:19 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____