
Midmark M11 - v1.0.5
Total Cycles: 3667

Name: RITTER2
Sterilizer ID: V2313228
06/04/2024 01:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.8 F 0.0
2:00 119.0 F 1.0
4:00 133.2 F 3.3
6:00 157.1 F 6.7
8:00 182.5 F 10.6
10:00 213.5 F 4.5
12:00 220.1 F 7.8
14:00 228.9 F 12.3
16:00 232.3 F 8.7
18:00 240.9 F 11.4
20:00 243.9 F 12.8
22:00 255.4 F 18.9
24:00 266.2 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.7 F 29.8
2:00 271.5 F 29.0
3:00 272.1 F 29.8
4:00 272.5 F 29.4

Min 271.0 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.9 F 0.0
5:00 243.7 F 0.0
10:00 224.1 F 0.0
15:00 228.8 F 0.0
20:00 227.0 F 0.0
25:00 230.9 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 25:44
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:04:00
CYCLE COMPLETE

06/04/2024 02:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____