
Midmark M11 - v1.0.5
Total Cycles: 2533

Name: RITTER2
Sterilizer ID: V2313228
06/19/2023 03:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.7 F 0.0
2:00 180.8 F 3.9
4:00 195.2 F 8.7
6:00 212.8 F 9.4
8:00 229.2 F 9.4
10:00 236.4 F 9.3
12:00 243.7 F 13.1
14:00 243.1 F 12.2
16:00 255.7 F 18.8
18:00 266.6 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.3 F 29.6
2:00 271.3 F 29.1
3:00 272.4 F 29.7
4:00 271.5 F 29.1

Min 271.2 F 29.0
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.2 F 0.0
5:00 211.0 F 0.0
10:00 219.1 F 0.0
15:00 227.9 F 0.0
20:00 227.3 F 0.0
25:00 224.8 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 2:14
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

06/19/2023 04:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____