
Midmark M11 - v1.0.5
Total Cycles: 4427

Name: RITTER2
Sterilizer ID: V2313228
03/13/2025 04:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.7 F 0.0
2:00 174.0 F 2.9
4:00 177.7 F 7.5
6:00 198.7 F 12.6
8:00 219.8 F 6.8
10:00 229.6 F 11.7
12:00 233.1 F 8.8
14:00 240.0 F 10.4
16:00 246.1 F 13.8
18:00 258.0 F 20.2
20:00 268.5 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 272.7 F 29.7
1:00 272.5 F 29.8
2:00 272.3 F 29.8
3:00 272.4 F 29.9
4:00 272.3 F 30.1

Min 271.9 F 29.6
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 225.9 F 0.0
10:00 221.2 F 0.0
15:00 198.3 F 0.0
20:00 192.7 F 0.0
25:00 189.9 F 0.0
30:00 188.1 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 21:50
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:00:57
CYCLE COMPLETE

03/13/2025 05:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____