
Midmark M11 - v1.0.0
Total Cycles: 3869

Name: RITTER 1
Sterilizer ID: V2178198
10/05/2023 08:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.5 F 0.0
2:00 176.2 F 2.2
4:00 177.3 F 5.6
6:00 201.5 F 10.1
8:00 221.1 F 5.0
10:00 225.1 F 9.2
12:00 238.6 F 10.9
14:00 240.1 F 11.3
16:00 238.0 F 9.5
18:00 251.0 F 15.9
20:00 262.2 F 22.3
22:00 272.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.8 F 29.1
2:00 271.4 F 28.4
3:00 272.8 F 29.2
4:00 271.5 F 28.4

Min 271.3 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 230.7 F 0.0
10:00 225.2 F 0.0
15:00 227.6 F 0.0
20:00 229.3 F 0.0
25:00 225.0 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 22:25
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:01:29
CYCLE COMPLETE

10/05/2023 09:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____