

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.0uS / 0.6ppm  
CYCLE NUMBER 001607  
14:16 11/22/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:47  
PRESSURIZING 6:32  
STERILIZING 7:58  
136.0 C 220kPa 7:58  
136.0 C 218kPa 8:28  
136.4 C 223kPa 8:58  
136.3 C 221kPa 9:28  
136.4 C 223kPa 9:58  
136.4 C 223kPa 10:28  
136.2 C 220kPa 10:58  
136.4 C 220kPa 11:28  
VENTING 11:28  
AIR DRYING 12:33  
CYCLE COMPLETE 37:51

-----  
Digital Signature #  
10D90C6A6DA4B7E9  
-----