

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002577  
11:49 14/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 2:40  
STERILIZING 3:44  
136.5 C 225kPa 3:44  
136.2 C 224kPa 4:14  
136.3 C 224kPa 4:44  
136.1 C 223kPa 5:14  
136.1 C 222kPa 5:44  
136.4 C 225kPa 6:14  
136.2 C 224kPa 6:44  
136.4 C 225kPa 7:14  
136.4 C 225kPa 7:14  
VENTING 7:14  
AIR DRYING 8:10  
CYCLE COMPLETE 38:10  
-----

Digital Signature #  
F6671EEB248220AA  
-----