
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002527
15:14 20/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 2:00
STERILIZING 3:01
136.4 C 225kPa 3:01
136.1 C 223kPa 3:31
136.4 C 226kPa 4:01
136.1 C 223kPa 4:31
136.0 C 222kPa 5:01
136.4 C 226kPa 5:31
136.4 C 225kPa 6:01
136.4 C 226kPa 6:31
136.4 C 226kPa 6:31
VENTING 6:31
AIR DRYING 7:24
CYCLE COMPLETE 19:02

Digital Signature #
B44A038E6BE24857
