

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1938

Name: RITTER2  
Sterilizer ID: V2313228  
01/17/2023 01:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.9 F 0.0  
2:00 172.0 F 3.9  
4:00 186.6 F 8.2  
6:00 207.8 F 11.3  
8:00 224.6 F 9.0  
10:00 232.2 F 9.6  
12:00 240.2 F 12.3  
14:00 240.2 F 11.1  
16:00 253.7 F 18.0  
18:00 265.8 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.4  
1:00 272.4 F 30.2  
2:00 271.9 F 29.4  
3:00 271.8 F 29.9  
4:00 272.4 F 29.9  
  
Min 271.2 F 29.2  
Max 272.6 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.6 F 0.0  
5:00 221.2 F 0.0  
10:00 232.3 F 0.0  
15:00 230.0 F 0.0  
20:00 237.0 F 0.0  
25:00 228.4 F 0.0  
30:00 231.1 F 0.0  
  
Drying Complete

FILLING: 1:21  
HEATING: 19:31  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:04  
CYCLE COMPLETE

01/17/2023 02:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_