
Midmark M11 - v1.0.5
Total Cycles: 3057

Name: RITTER2
Sterilizer ID: V2313228
10/27/2023 10:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.8 F 0.0
2:00 167.5 F 2.8
4:00 180.6 F 6.8
6:00 201.6 F 11.8
8:00 218.6 F 6.2
10:00 231.8 F 11.6
12:00 233.9 F 9.0
14:00 238.2 F 9.6
16:00 248.3 F 15.0
18:00 260.8 F 22.0
20:00 272.0 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.6 F 29.6
2:00 271.5 F 29.1
3:00 272.7 F 29.7
4:00 271.7 F 29.0

Min 271.5 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.0 F 0.0
5:00 210.9 F 0.0
10:00 220.1 F 0.0
15:00 214.0 F 0.0
20:00 210.6 F 0.0
25:00 209.5 F 0.0
30:00 207.8 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 20:32
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

10/27/2023 11:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____