
Midmark M11 - v1.0.5
Total Cycles: 2582

Name: RITTER2
Sterilizer ID: V2313228
06/28/2023 01:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.2 F 0.0
2:00 173.8 F 3.9
4:00 184.8 F 7.9
6:00 203.6 F 12.6
8:00 218.7 F 6.4
10:00 231.8 F 11.3
12:00 233.1 F 8.5
14:00 240.3 F 10.9
16:00 246.2 F 13.8
18:00 258.6 F 20.6
20:00 269.8 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.4 F 29.4
2:00 272.3 F 29.9
3:00 271.8 F 29.1
4:00 272.6 F 29.8

Min 271.5 F 29.0
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 207.9 F 0.0
10:00 212.7 F 0.0
15:00 213.5 F 0.0
20:00 210.5 F 0.0
25:00 206.9 F 0.0
30:00 204.3 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:55
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:59:33
CYCLE COMPLETE

06/28/2023 02:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____