

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4328

Name: RITTER 1  
Sterilizer ID: V2178198  
02/23/2024 02:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 98.7 F 0.0  
2:00 100.1 F 0.8  
4:00 117.1 F 3.0  
6:00 147.3 F 5.8  
8:00 170.8 F 9.1  
10:00 189.4 F 12.8  
12:00 204.6 F 5.1  
14:00 215.1 F 9.0  
16:00 227.0 F 13.1  
18:00 232.9 F 9.0  
20:00 240.3 F 11.5  
22:00 242.4 F 12.0  
24:00 253.8 F 17.8  
26:00 264.2 F 23.9  
28:00 271.8 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.9  
1:00 272.2 F 29.5  
2:00 271.9 F 29.0  
3:00 271.8 F 29.3  
4:00 272.5 F 29.3

Min 271.2 F 28.7  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.2 F 0.0  
5:00 205.3 F 0.0  
10:00 204.9 F 0.0  
15:00 206.5 F 0.0  
20:00 207.5 F 0.0  
25:00 209.5 F 0.0  
30:00 210.4 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 28:01  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:06:36  
CYCLE COMPLETE

02/23/2024 03:23 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_