
Midmark M11 - v1.0.5
Total Cycles: 4041

Name: RITTER2
Sterilizer ID: V2313228
10/18/2024 08:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 84.8 F 0.0
4:00 79.2 F 0.0
6:00 108.3 F 0.0
8:00 144.0 F 0.2
10:00 172.1 F 0.4
12:00 195.6 F 2.7
14:00 218.0 F 6.2
16:00 227.3 F 9.8
18:00 234.1 F 11.8
20:00 234.3 F 9.2
22:00 241.3 F 11.6
24:00 242.0 F 11.7
26:00 239.6 F 10.2
28:00 252.4 F 16.6
30:00 263.3 F 23.0
32:00 272.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.5 F 29.0
2:00 271.5 F 28.4
3:00 272.6 F 29.1
4:00 271.6 F 28.3

Min 271.3 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.6 F 0.0
5:00 231.6 F 0.0
10:00 220.8 F 0.0
15:00 194.9 F 0.0
20:00 198.5 F 0.0
25:00 203.0 F 0.0
30:00 197.7 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:16
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:10:24
CYCLE COMPLETE

10/18/2024 09:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____