
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000225
8:49 09/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:57
STERILIZING 3:54
136.0 C 220kPa 3:54
136.1 C 222kPa 4:24
136.2 C 223kPa 4:54
136.4 C 224kPa 5:24
136.4 C 224kPa 5:54
136.1 C 221kPa 6:24
136.4 C 224kPa 6:54
136.5 C 224kPa 7:24
136.5 C 224kPa 7:24
VENTING 7:24
AIR DRYING 8:12
CYCLE COMPLETE 8:15

Digital Signature #
17981D65EF2CC09C
