
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001246
15:24 23/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:40
PRESSURIZING 1:45
STERILIZING 2:53
136.3 C 223kPa 2:53
136.4 C 223kPa 3:23
136.4 C 225kPa 3:53
136.1 C 219kPa 4:23
136.3 C 222kPa 4:53
136.4 C 222kPa 5:23
136.4 C 225kPa 5:53
136.4 C 224kPa 6:23
136.4 C 224kPa 6:23
VENTING 6:23
AIR DRYING 7:19
CYCLE COMPLETE 7:19

Digital Signature #
0D07D039FF87F7BD
