

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008192  
8:41 08/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:52  
PRESSURIZING 2:54  
STERILIZING 3:56  
136.4 C 221kPa 3:56  
136.0 C 220kPa 4:26  
136.4 C 221kPa 4:56  
136.4 C 219kPa 5:26  
136.1 C 217kPa 5:56  
136.2 C 219kPa 6:26  
136.4 C 221kPa 6:56  
136.3 C 222kPa 7:26  
136.3 C 222kPa 7:26  
VENTING 7:26  
AIR DRYING 8:15  
CYCLE COMPLETE 25:21

-----  
Digital Signature #  
D6FAD046D842780A  
-----