
Midmark M11 - v1.0.5
Total Cycles: 3116

Name: RITTER2
Sterilizer ID: V2313228
11/13/2023 12:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.4 F 0.0
2:00 167.7 F 2.2
4:00 171.1 F 6.0
6:00 192.6 F 10.7
8:00 210.6 F 5.0
10:00 227.2 F 9.7
12:00 230.7 F 9.1
14:00 242.4 F 12.9
16:00 242.7 F 12.0
18:00 255.5 F 18.7
20:00 266.9 F 25.8
22:00 272.4 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.3 F 29.5
2:00 272.9 F 29.9
3:00 271.6 F 29.7
4:00 272.0 F 29.5

Min 271.3 F 29.3
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.2 F 0.0
5:00 204.2 F 0.0
10:00 197.4 F 0.0
15:00 195.3 F 0.0
20:00 195.3 F 0.0
25:00 194.6 F 0.0
30:00 194.3 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 22:08
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:00:44
CYCLE COMPLETE

11/13/2023 01:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____