
Midmark M11 - v1.0.5
Total Cycles: 2590

Name: RITTER2
Sterilizer ID: V2313228
06/29/2023 02:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.6 F 0.0
2:00 166.7 F 2.8
4:00 180.6 F 6.9
6:00 201.9 F 11.8

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 209F
CHAMBER PRESS 11.6PSI
** ITEMS NOT STERILE **

06/29/2023 03:02 PM