
Midmark M11 - v1.0.5
Total Cycles: 1581

Name: RITTER2
Sterilizer ID: V2313228
10/12/2022 07:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.2 F 0.0

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 79F
CHAMBER PRESS 0.0PSI

** ITEMS NOT STERILE **

10/12/2022 07:25 AM