
Midmark M11 - v1.0.5
Total Cycles: 3907

Name: RITTER2
Sterilizer ID: V2313228
09/05/2024 08:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.1 F 0.0
2:00 70.7 F 0.0
4:00 75.8 F 0.0
6:00 100.3 F 0.1
8:00 139.7 F 1.1
10:00 169.6 F 3.9
12:00 190.2 F 7.2
14:00 203.6 F 11.4
16:00 215.7 F 4.5
18:00 226.3 F 8.7
20:00 236.4 F 13.1
22:00 236.5 F 9.9
24:00 237.9 F 9.3
26:00 247.0 F 14.0
28:00 258.1 F 20.1
30:00 267.8 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.3
2:00 271.7 F 29.1
3:00 272.2 F 29.0
4:00 271.9 F 29.3

Min 271.5 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.0 F 0.0
5:00 204.9 F 0.0
10:00 194.2 F 0.0
15:00 191.3 F 0.0
20:00 193.8 F 0.0
25:00 197.3 F 0.0
30:00 204.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:24
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:29
CYCLE COMPLETE

09/05/2024 09:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____