
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004822
15:53 06/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:03
STERILIZING 4:18
136.4 C 222kPa 4:18
136.4 C 225kPa 4:48
136.4 C 225kPa 5:18
136.4 C 224kPa 5:48
136.4 C 225kPa 6:18
136.4 C 225kPa 6:48
136.4 C 225kPa 7:18
136.3 C 225kPa 7:48
136.3 C 225kPa 7:48
VENTING 7:48
AIR DRYING 8:40
CYCLE COMPLETE 17:31

Digital Signature #
37668D88EFB943AD
