
Midmark M11 - v1.0.5
Total Cycles: 4505

Name: RITTER2
Sterilizer ID: V2313228
04/11/2025 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 71.1 F 0.0
4:00 79.2 F 0.4
6:00 123.4 F 2.5
8:00 141.3 F 5.5
10:00 168.2 F 9.1
12:00 193.2 F 12.7
14:00 204.3 F 4.2
16:00 220.9 F 7.5
18:00 230.9 F 11.3
20:00 225.9 F 6.2
22:00 238.8 F 11.2
24:00 235.2 F 8.5
26:00 247.6 F 14.5
28:00 258.2 F 20.3
30:00 268.1 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.6 F 29.3
2:00 272.4 F 29.7
3:00 271.9 F 28.9
4:00 272.9 F 29.6

Min 271.6 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 203.8 F 0.0
10:00 209.0 F 0.0
15:00 214.8 F 0.0
20:00 219.5 F 0.0
25:00 224.5 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:18
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:09:37
CYCLE COMPLETE

04/11/2025 09:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____