
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002141
12:41 10/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:06
STERILIZING 3:05
136.5 C 223kPa 3:05
136.0 C 223kPa 3:35
136.4 C 222kPa 4:05
136.2 C 223kPa 4:35
136.3 C 224kPa 5:05
136.4 C 224kPa 5:35
136.4 C 225kPa 6:05
136.2 C 224kPa 6:35
136.2 C 224kPa 6:35
VENTING 6:35
AIR DRYING 7:25
CYCLE COMPLETE 37:25

Digital Signature #
452C12FAF0FE29A4
