

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1999

Name: RITTER2  
Sterilizer ID: V2313228  
02/02/2023 11:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.1 F 0.0  
2:00 169.9 F 2.0  
4:00 192.2 F 6.0  
6:00 212.6 F 11.3  
8:00 220.8 F 6.2  
10:00 232.1 F 11.7  
12:00 234.3 F 9.0  
14:00 237.9 F 9.8  
16:00 247.5 F 14.5  
18:00 259.9 F 21.5  
20:00 271.0 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.4 F 29.9  
2:00 271.5 F 29.1  
3:00 272.3 F 30.0  
4:00 271.9 F 29.5

Min 271.4 F 29.1  
Max 272.4 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.8 F 0.0  
5:00 214.8 F 0.0  
10:00 226.1 F 0.0  
15:00 227.6 F 0.0  
20:00 228.9 F 0.0  
25:00 229.7 F 0.0  
30:00 230.0 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 20:43  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:58  
CYCLE COMPLETE

02/02/2023 12:25 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_