
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000560
15:32 29/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:28
PRESSURIZING 4:28
STERILIZING 5:46
136.0 C 228kPa 5:46
136.2 C 231kPa 6:16
136.2 C 231kPa 6:46
136.2 C 230kPa 7:16
136.5 C 230kPa 7:46
136.2 C 230kPa 8:16
136.3 C 232kPa 8:46
136.4 C 232kPa 9:16
136.4 C 232kPa 9:16
VENTING 9:17
AIR DRYING 10:14
CYCLE COMPLETE 25:14

Digital Signature #
6744667E419ED232
