
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000259
13:08 30/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:17
PRESSURIZING 3:26
STERILIZING 4:57
136.5 C 232kPa 4:57
136.2 C 231kPa 5:27
136.4 C 233kPa 5:57
136.4 C 233kPa 6:27
136.4 C 233kPa 6:57
136.4 C 233kPa 7:27
136.3 C 232kPa 7:57
136.4 C 233kPa 8:27
136.4 C 233kPa 8:27
VENTING 8:27
AIR DRYING 9:35
CYCLE COMPLETE 24:35

Digital Signature #
9D43261AA14067D0
