

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4761

Name: RITTER 1  
Sterilizer ID: V2178198  
07/24/2024 08:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.2 F 0.0  
2:00 77.2 F 0.0  
4:00 87.2 F 0.1  
6:00 120.0 F 1.4  
8:00 152.1 F 3.8  
10:00 177.0 F 6.7  
12:00 191.1 F 10.1  
14:00 199.0 F 8.9  
16:00 212.4 F 5.1  
18:00 225.3 F 8.7  
20:00 237.6 F 12.8  
22:00 234.1 F 8.8  
24:00 242.6 F 11.3  
26:00 244.6 F 12.5  
28:00 256.6 F 18.8  
30:00 267.3 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.7 F 28.7  
2:00 271.8 F 28.4  
3:00 272.6 F 28.7  
4:00 271.8 F 28.3

Min 271.6 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
5:00 203.9 F 0.0  
10:00 215.6 F 0.0  
15:00 216.4 F 0.0  
20:00 215.5 F 0.0  
25:00 220.7 F 0.0  
30:00 222.9 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:26  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:09:40  
CYCLE COMPLETE

07/24/2024 10:01 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_