
Midmark M11 - v1.0.5
Total Cycles: 3659

Name: RITTER2
Sterilizer ID: V2313228
05/30/2024 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.6 F 0.0
2:00 106.6 F 0.2
4:00 121.5 F 1.6
6:00 149.6 F 4.4
8:00 173.1 F 8.1
10:00 193.1 F 12.1
12:00 211.4 F 5.1
14:00 224.0 F 9.2
16:00 233.9 F 12.1
18:00 235.1 F 10.4
20:00 235.2 F 8.4
22:00 246.9 F 14.3
24:00 257.9 F 20.4
26:00 268.1 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.3 F 30.1
2:00 272.5 F 29.6
3:00 271.2 F 29.2
4:00 272.0 F 30.0

Min 271.2 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 208.2 F 0.0
10:00 193.7 F 0.0
15:00 192.5 F 0.0
20:00 188.7 F 0.0
25:00 186.3 F 0.0
30:00 192.2 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 27:26
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:05:38
CYCLE COMPLETE

05/30/2024 11:21 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____