
Midmark M11 - v1.0.0
Total Cycles: 2682

Name: RITTER 1
Sterilizer ID: V2178198
12/15/2022 10:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.1 F 0.0
2:00 196.9 F 3.0
4:00 202.8 F 7.5
6:00 226.3 F 12.7
8:00 229.7 F 8.7
10:00 239.8 F 10.7
12:00 241.0 F 12.0
14:00 240.0 F 10.5
16:00 253.7 F 17.4
18:00 265.5 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.5 F 29.2
2:00 271.9 F 28.5
3:00 272.1 F 29.1
4:00 272.4 F 29.1

Min 271.5 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 204.0 F 0.0
10:00 212.4 F 0.0
15:00 220.6 F 0.0
20:00 223.5 F 0.0
25:00 224.3 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 19:46
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:06
CYCLE COMPLETE

12/15/2022 11:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____