
Midmark M11 - v1.0.5
Total Cycles: 3175

Name: RITTER2
Sterilizer ID: V2313228
12/01/2023 11:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.8 F 0.0
2:00 157.0 F 0.5
4:00 152.7 F 3.0
6:00 174.8 F 6.6
8:00 190.8 F 10.8
10:00 215.3 F 5.2
12:00 224.0 F 9.0
14:00 234.6 F 11.4
16:00 237.6 F 11.2
18:00 237.0 F 9.6
20:00 249.8 F 15.9
22:00 261.0 F 22.3
24:00 271.4 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.3
1:00 272.7 F 30.1
2:00 271.9 F 29.5
3:00 271.2 F 29.2
4:00 272.2 F 30.1

Min 271.1 F 29.1
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 226.1 F 0.0
10:00 226.3 F 0.0
15:00 235.6 F 0.0
20:00 225.8 F 0.0
25:00 221.6 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 24:44
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:03:19
CYCLE COMPLETE

12/01/2023 12:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____