
Midmark M11 - v1.0.5
Total Cycles: 3524

Name: RITTER2
Sterilizer ID: V2313228
04/12/2024 03:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.1 F 0.0
2:00 174.3 F 3.2
4:00 190.1 F 7.7
6:00 209.0 F 12.6
8:00 218.1 F 7.2
10:00 231.4 F 12.2
12:00 234.9 F 9.6
14:00 236.9 F 9.5
16:00 247.7 F 14.9
18:00 259.8 F 21.5
20:00 270.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.0 F 29.6
2:00 272.2 F 30.0
3:00 271.6 F 29.3
4:00 272.4 F 30.0

Min 271.1 F 29.2
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.9 F 0.0
5:00 233.6 F 0.0
10:00 200.0 F 0.0
15:00 190.4 F 0.0
20:00 186.5 F 0.0
25:00 184.4 F 0.0
30:00 183.2 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:46
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 00:59:58
CYCLE COMPLETE

04/12/2024 04:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____