

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007236  
9:17 22/04/2024

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:00  
PRESSURIZING 3:02  
STERILIZING 4:06  
136.0 C 220kPa 4:06  
136.5 C 223kPa 4:36  
136.4 C 224kPa 5:06  
136.5 C 224kPa 5:36  
136.1 C 219kPa 6:06  
136.3 C 224kPa 6:36  
136.4 C 224kPa 7:06  
136.1 C 220kPa 7:36  
136.1 C 220kPa 7:36  
VENTING 7:36  
AIR DRYING 8:28  
CYCLE COMPLETE 31:26  
-----

Digital Signature #  
D3E6931FDD559EFC  
-----