
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006674
14:47 29/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:48
PRESSURIZING 1:54
STERILIZING 2:47
136.0 C 220kPa 2:47
136.3 C 222kPa 3:17
136.4 C 223kPa 3:47
136.3 C 224kPa 4:17
136.1 C 219kPa 4:47
136.2 C 223kPa 5:17
136.4 C 220kPa 5:47
136.4 C 221kPa 6:17
136.4 C 221kPa 6:17
VENTING 6:17
AIR DRYING 7:08
CYCLE COMPLETE 7:14

Digital Signature #
FC41FE1B25D6DC3D
