
Midmark M11 - v1.0.0
Total Cycles: 3439

Name: RITTER 1
Sterilizer ID: V2178198
06/23/2023 03:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.8 F 0.0
2:00 191.2 F 3.8
4:00 212.1 F 9.1
6:00 227.6 F 8.0
8:00 237.3 F 11.0
10:00 235.0 F 8.5
12:00 239.8 F 10.0
14:00 248.6 F 14.4
16:00 261.3 F 21.5
18:00 272.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.4 F 28.7
2:00 271.4 F 28.2
3:00 272.4 F 28.8
4:00 271.5 F 28.1

Min 271.3 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.0 F 0.0
5:00 236.9 F 0.0
10:00 222.6 F 0.0
15:00 202.9 F 0.0
20:00 193.2 F 0.0
25:00 189.5 F 0.0
30:00 191.9 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 18:25
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

06/23/2023 04:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____