
Midmark M11 - v1.0.0
Total Cycles: 4522

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2024 10:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.4 F 0.0
2:00 139.7 F 2.0
4:00 161.0 F 4.9
6:00 188.8 F 8.8
8:00 201.2 F 12.8
10:00 217.3 F 5.5
12:00 225.5 F 9.5
14:00 237.6 F 10.8
16:00 237.8 F 10.2
18:00 236.6 F 8.7
20:00 247.5 F 13.8
22:00 258.1 F 19.7
24:00 268.2 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 271.9 F 29.0
2:00 272.1 F 28.6
3:00 272.6 F 29.2
4:00 271.5 F 28.5

Min 270.8 F 28.0
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.4 F 0.0
5:00 234.1 F 0.0
10:00 192.2 F 0.0
15:00 186.8 F 0.0
20:00 185.8 F 0.0
25:00 186.2 F 0.0
30:00 203.9 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 25:24
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:04:14
CYCLE COMPLETE

05/03/2024 12:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____