
Midmark M11 - v1.0.0
Total Cycles: 5009

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2024 10:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.9 F 0.0
2:00 102.6 F 0.0
4:00 125.4 F 0.7
6:00 153.6 F 1.4
8:00 180.7 F 4.5
10:00 202.7 F 7.7
12:00 211.6 F 11.5
14:00 217.2 F 4.8
16:00 229.2 F 9.1
18:00 238.8 F 12.2
20:00 239.2 F 10.5
22:00 236.1 F 8.1
24:00 249.1 F 14.5
26:00 260.2 F 20.6
28:00 270.5 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.6 F 28.7
2:00 271.9 F 28.1
3:00 271.7 F 28.2
4:00 272.6 F 28.7

Min 271.4 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.1 F 0.0
5:00 202.1 F 0.0
10:00 196.1 F 0.0
15:00 187.4 F 0.0
20:00 185.2 F 0.0
25:00 184.2 F 0.0
30:00 184.2 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 28:53
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:07:18
CYCLE COMPLETE

10/18/2024 11:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____