
Midmark M11 - v1.0.5
Total Cycles: 4134

Name: RITTER2
Sterilizer ID: V2313228
11/19/2024 09:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 70.3 F 0.0
4:00 78.5 F 0.0
6:00 105.2 F 0.1
8:00 145.4 F 0.4
10:00 173.5 F 0.7
12:00 194.9 F 2.5
14:00 213.4 F 5.8
16:00 219.9 F 9.1
18:00 237.5 F 12.8
20:00 232.3 F 8.3
22:00 241.1 F 13.0
24:00 240.0 F 10.8
26:00 236.9 F 9.2
28:00 250.5 F 15.7
30:00 261.5 F 21.9
32:00 271.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.4 F 29.2
2:00 271.2 F 28.5
3:00 272.6 F 29.2
4:00 271.4 F 28.4

Min 271.2 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.4 F 0.0
5:00 238.2 F 0.0
10:00 227.5 F 0.0
15:00 207.1 F 0.0
20:00 180.7 F 0.0
25:00 173.0 F 0.0
30:00 171.3 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 32:38
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:10:48
CYCLE COMPLETE

11/19/2024 11:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____