

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000469  
14:06 02/02/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 2:39  
STERILIZING 3:47  
136.0 C 228kPa 3:47  
136.2 C 231kPa 4:17  
136.4 C 233kPa 4:47  
136.2 C 231kPa 5:17  
136.4 C 233kPa 5:47  
136.1 C 231kPa 6:17  
136.4 C 233kPa 6:47  
136.4 C 233kPa 7:17  
136.4 C 233kPa 7:17  
VENTING 7:18  
AIR DRYING 8:11  
CYCLE COMPLETE 8:14  
-----

Digital Signature #  
9009ECC3CC4C281C  
-----