
Midmark M11 - v1.0.0
Total Cycles: 776

Name: RITTER 1
Sterilizer ID: V2178198
04/07/2021 01:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.2 F 0.0
2:00 198.2 F 3.4
4:00 216.1 F 8.0
6:00 228.6 F 9.5
8:00 231.0 F 9.7
10:00 235.5 F 8.1
12:00 244.9 F 13.5
14:00 244.1 F 12.2
16:00 257.0 F 19.1
18:00 268.4 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.2 F 28.8
2:00 271.7 F 29.0
3:00 271.8 F 28.7
4:00 271.9 F 29.1

Min 271.2 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 235.6 F 0.0
10:00 228.3 F 0.0
15:00 232.0 F 0.0
20:00 228.5 F 0.0
25:00 229.2 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 19:07
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:57:29
CYCLE COMPLETE

04/07/2021 02:19 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____