

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1324

Name: RITTER 1  
Sterilizer ID: V2178198  
11/01/2021 06:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.9 F 0.0  
2:00 77.4 F 0.0  
4:00 85.0 F 0.0  
6:00 147.2 F 0.0  
8:00 159.0 F 0.1  
10:00 198.2 F 1.3  
12:00 211.1 F 4.1  
14:00 219.5 F 6.8  
16:00 225.4 F 10.0  
18:00 233.0 F 13.5  
20:00 231.5 F 8.2  
22:00 241.4 F 12.5  
24:00 237.5 F 9.7  
26:00 239.1 F 9.2  
28:00 247.7 F 14.1  
30:00 259.0 F 20.3  
32:00 269.3 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.5 F 28.9  
2:00 271.6 F 28.5  
3:00 272.5 F 28.9  
4:00 271.5 F 28.3  
  
Min 271.4 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.9 F 0.0  
5:00 215.2 F 0.0  
10:00 217.1 F 0.0  
15:00 216.1 F 0.0  
20:00 214.2 F 0.0  
25:00 210.6 F 0.0  
30:00 198.7 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 33:03  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:11:14  
CYCLE COMPLETE

11/01/2021 07:59 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01