
Midmark M11 - v1.0.5
Total Cycles: 1512

Name: RITTER2
Sterilizer ID: V2313228
09/21/2022 08:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.3 F 0.0
2:00 167.7 F 2.3
4:00 178.2 F 6.1
6:00 197.8 F 11.1
8:00 220.4 F 6.2
10:00 232.0 F 11.8
12:00 234.4 F 9.1
14:00 239.3 F 9.5
16:00 247.2 F 14.5
18:00 259.3 F 21.4
20:00 270.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.6
1:00 272.6 F 30.2
2:00 271.7 F 29.5
3:00 272.1 F 30.3
4:00 272.5 F 30.1

Min 271.2 F 29.4
Max 272.7 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 228.4 F 0.0
10:00 230.1 F 0.0
15:00 227.9 F 0.0
20:00 225.0 F 0.0
25:00 227.8 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 20:51
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:10
CYCLE COMPLETE

09/21/2022 09:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____