
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000266
6:50 01/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:59
PRESSURIZING 4:01
STERILIZING 5:34
136.0 C 227kPa 5:34
136.2 C 231kPa 6:04
136.2 C 231kPa 6:34
136.4 C 232kPa 7:04
136.4 C 232kPa 7:34
136.4 C 233kPa 8:04
136.2 C 230kPa 8:34
136.4 C 231kPa 9:04
136.4 C 231kPa 9:04
VENTING 9:04
AIR DRYING 10:03
CYCLE COMPLETE 25:03

Digital Signature #
AD14ACB0BC4C2CB8
