
Midmark M11 - v1.0.5
Total Cycles: 2885

Name: RITTER2
Sterilizer ID: V2313228
09/14/2023 02:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.9 F 0.0
2:00 165.0 F 3.2
4:00 187.1 F 7.4
6:00 205.7 F 12.1
8:00 220.1 F 7.1
10:00 235.3 F 12.4
12:00 237.5 F 10.0
14:00 237.3 F 9.2
16:00 249.4 F 15.1
18:00 260.6 F 21.6
20:00 271.0 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 272.0 F 29.3
2:00 272.8 F 29.6
3:00 272.5 F 29.9
4:00 271.3 F 29.3

Min 271.3 F 29.2
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 243.5 F 0.0
10:00 191.3 F 0.0
15:00 190.2 F 0.0
20:00 187.1 F 0.0
25:00 184.7 F 0.0
30:00 183.2 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 21:18
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:00:51
CYCLE COMPLETE

09/14/2023 03:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____