
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
1.9uS / 1.2ppm
CYCLE NUMBER 000039
16:45 10/08/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:26
PRESSURIZING 2:34
STERILIZING 3:30
136.0 C 221kPa 3:30
136.3 C 227kPa 4:00
136.5 C 226kPa 4:30
136.1 C 224kPa 5:00
136.4 C 224kPa 5:31
136.3 C 226kPa 6:01
136.3 C 227kPa 6:31
136.4 C 227kPa 7:00
VENTING 7:00
AIR DRYING 7:52
CYCLE COMPLETE 9:52

Digital Signature #
797D4E5D0E6EE89B
